



VITROMED
Germany

VITROMED GmbH
Raiffeisenstraße 15a
40764 Langenfeld
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 010VM3401

PRODUCT NAME: V-PVP 7%

REFERENCE: V-PVP-7

EXPIRY DATE: 28.10.2021

TEST	SPECIFICATION	RESULT
pH	7,25 -7,45	7,41
OSMOLALITY	300 – 330 mOsm/kg	326
BACTERIAL SCREEN	< 0,5 EU/ml	PASSED
FUNGAL SCREEN	≥ 80 %	PASSED
ENDOTOXIN	PASSED	PASSED
MEA 1-cell to expanded blastocyst 96h	PASSED	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED