



**VITROMED**  
Germany

VITROMED GmbH  
Raiffeisenstraße 15a  
40764 Langenfeld  
GERMANY

Tel +49 2173-20041-30

[www.vitromed.com](http://www.vitromed.com)

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 011M1301

PRODUCT NAME: V-ONESTEP Culture Medium

REFERENCE: V-OSM-10

EXPIRY DATE: 07-Feb-2021

TEST	SPECIFICATION	RESULT
pH	7.25 – 7.40	7.34
OSMOLALITY	257 - 273	262
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocysts 120h	≥ 80%	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED