

CERTIFICATE OF ANALYSIS

LOT NUMBER: 011VM0501
PRODUCT NAME: V-SPERM WASH
REFERENCE: V-SWM-100
EXPIRY DATE: 18.11.2021

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7,25
OSMOLALITY	257-273 mOsm/kg	258
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

