

CERTIFICATE OF ANALYSIS

LOT NUMBER: 012VM2301

PRODUCT NAME: V-HEPES PLUS 100

REFERENCE: V-HTF-P1

EXPIRY DATE: 08.12.2021

| TEST | SPECIFICATION | RESULT |
|---|-----------------|--------|
| pH | 7.00-7.40 | 7.21 |
| OSMOLALITY | 257-273 mOsm/kg | 261 |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| ENDOTOXIN | < 0.25 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |
| Color | PASSED | PASSED |