



**VITROMED**  
Germany

VITROMED GmbH  
Raiffeisenstraße 15a  
40764 Langenfeld  
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 102VM2001  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-20P  
EXPIRY DATE: 05-May-2021

TEST	SPECIFICATION	RESULT
pH	7.20 – 7.40	7.28
OSMOLALITY	257 - 273	264
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocysts 120h	≥ 80%	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED