



VITROMED
Germany

VITROMED GmbH
Raiffeisenstraße 15a
40764 Langenfeld
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103M1302
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-10
EXPIRY DATE: 22-Jun-2021

R_x Only Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

| TEST | SPECIFICATION | RESULT |
|--|--------------------------|-------------|
| pH | 7.20 – 7.40 | 7.31 |
| OSMOLALITY | 257 – 273 mOsm/kg | 263 mOsm/kg |
| STERILITY: Sterile (no growth) | | |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| ENDOTOXIN | <0.25 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst 120h | ≥ 80% | PASSED |
| Color | Pink Rose to Pink Orange | PASSED |