



## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 103M1401  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-20  
EXPIRY DATE: 08-Jun-2021

**R<sub>x</sub> Only** Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

| TEST                                   | SPECIFICATION            | RESULT      |
|--|--------------------------|-------------|
| pH                                     | 7.20 – 7.40              | 7.28        |
| OSMOLALITY                             | 257 – 273 mOsm/kg        | 261 mOsm/kg |
| <b>STERILITY: Sterile (no growth)</b>  |                          |             |
| BACTERIAL SCREEN                       | PASSED                   | PASSED      |
| FUNGAL SCREEN                          | PASSED                   | PASSED      |
| ENDOTOXIN                              | <0.25 EU/mL              | PASSED      |
| MEA 1-cell to expanded blastocyst 120h | ≥ 80%                    | PASSED      |
| Color                                  | Pink Rose to Pink Orange | PASSED      |