

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103M1402
 PRODUCT NAME: V-ONESTEP Culture Medium
 REFERENCE: V-OSM-20
 EXPIRY DATE: 22-Jun-2021

R_x Only Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

TEST	SPECIFICATION	RESULT
pH	7.20 – 7.40	7.31
OSMOLALITY	257 – 273 mOsm/kg	263 mOsm/kg
STERILITY: Sterile (no growth)		
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst 120h	≥ 80%	PASSED
Color	Pink Rose to Pink Orange	PASSED

