

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103VM0502
PRODUCT NAME: V-SPERM WASH
REFERENCE: V-SWM-100
EXPIRY DATE: 25-Mar-2022

R_x Only Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7,27
OSMOLALITY	257-273 mOsm/kg	263
STERILITY: Sterile (no growth)		
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

