

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103VM3601
PRODUCT NAME: V-HYADASE
REFERENCE: V-HYA
EXPIRY DATE: 24.02.2022

TEST	SPECIFICATION	RESULT
pH	7,20 – 7,40	7.25
OSMOLALITY	257 – 273 mOsm/kg	264
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,00 EU/ml	PASSED
MEA 1-cell to expanded blastocyst 96 h	≥ 80 %	PASSED

