

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 104VM1301  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-10  
EXPIRY DATE: 11-Jul-2021

**R<sub>x</sub> Only** Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

TEST	SPECIFICATION	RESULT
pH	7.20 – 7.40	7,22
OSMOLALITY	257 – 273 mOsm/kg	261 mOsm/kg
<b>STERILITY: Sterile (no growth)</b>		
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst 120h	≥ 80%	PASSED
Color	Pink Rose to Pink Orange	PASSED

