



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18100IN352
PRODUCT NAME: V-ICSIPET INJECTION S2 35°
REFERENCE: V-INJ-S2-35
EXPIRY DATE: 19.01.2021

| TEST | SPECIFICATION | RESULT |
|-------------------------------------------|---------------|--------|
| GAMMA STERILIZATION | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |

F. Buch