

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18118HO35  
PRODUCT NAME: V-ICSIPET HOLDING 35°  
REFERENCE: V-HOLD-35  
EXPIRY DATE: 12.02.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED