

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18120IN353
PRODUCT NAME: V-ICSIPET INJECTION S3 35°
REFERENCE: V-INJ-S3-35
EXPIRY DATE: 12.02.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED