



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18132HO30  
PRODUCT NAME: V-ICSIPET HOLDING 30°  
REFERENCE: V-HOLD-30  
EXPIRY DATE: 12.03.2021

| TEST                                      | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION                       | 25 - 45 kGy   | PASSED |
| ENDOTOXIN                                 | < 0,05 EU/mL  | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 %        | PASSED |

*F. Buch*