

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18134VD35
PRODUCT NAME: V-DENUPET 135
REFERENCE: V-DEN-135
EXPIRY DATE: 28.03.2021

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |

