



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18264HO25, 18265HO30, 18304HO35, 18303HO20, 18266IN303, 18309IN302

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-25/30/35/20, V-INJ-S3-30, V-INJ-S2-30

EXPIRY DATE: 28.05.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

F. Buch