



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18313IN303, 18339IN353, 18338IN302, 18340HO35, 18314HO30

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-30, V-INJ-S3-35, V-INJ-S2-30, V-HOLD-35, V-HOLD-30

EXPIRY DATE: 03.07.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED