



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18313IN303, 18339IN353, 18338IN302, 18340HO35, 18314HO30

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-30, V-INJ-S3-35, V-INJ-S2-30, V-HOLD-35, V-HOLD-30

EXPIRY DATE: 03.07.2021

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |