

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18361VD25
PRODUCT NAME: V-DENUPET 125
REFERENCE: V-DEN-125
EXPIRY DATE: 13.08.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

VITROMED GmbH

Wildenbruchstr. 15
07745 Jena - Germany
Tel +49 3641539 -1976, Fax -1977
office@vitromed.com
www.vitromed.com