

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18364VD50
PRODUCT NAME: V-DENUPET 150
REFERENCE: V-DEN-150
EXPIRY DATE: 13.08.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena - Germany
Tel +49 3641539 -1976, Fax -1977
office@vitromed.com
www.vitromed.com

i. A. 