

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18365VD75  
PRODUCT NAME: V-DENUPET 175  
REFERENCE: V-DEN-175  
EXPIRY DATE: 13.08.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

**VITROMED GmbH**  
Wildenbruchstr. 15  
07745 Jena - Germany  
Tel +49 3641539 -1976, Fax -1977  
office@vitromed.com  
www.vitromed.com