

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18368VD60
PRODUCT NAME: V-DENUPET 600
REFERENCE: V-DEN-600
EXPIRY DATE: 13.08.2021

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena - Germany
Tel +49 3641539 -1976, Fax -1977
office@vitromed.com
www.vitromed.com

[Handwritten signature]