



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18421HO35, 18420HO30, 18418IN303, 18419IN353

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-30, V-INJ-S3-35, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 12.09.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED