



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 509256
PRODUCT V-MARKER-PEN
Product Number V-PEN-4C
Expiry Date:

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0,03 EU/device	< 0,03 EU/mL
MEA 1-cell to expanded blactocyst in 96h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.12.2015