

CERTIFICATE OF ANALYSIS

Lot Number: 512GP15NP
PRODUCT V-GLASSPET 5.3/4 (150) NP
Product Number V-GPT-150NP
Expiry Date: 30.11.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96h	≥ 80 %	PASSED

PRODUCT RELEASED BY Franziska Buch ON 26.05.2015

