

## CERTIFICATE OF ANALYSIS

**Lot Number:** 512GP23  
**PRODUCT** V-GLASSPET 9 (230)  
Product Number V-GPT-230  
Expiry Date: 30.11.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96h	≥ 80 %	PASSED

PRODUCT RELEASED BY Franziska Buch ON 26.05.2015

