

CERTIFICATE OF ANALYSIS

Lot Number: 6050117
PRODUCT V-DENUPET 175
Product Number V-DEN-175
Expiry Date: 13.05.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 27.05.2016

