

CERTIFICATE OF ANALYSIS

Lot Number: 605M0401 V-GRAD 40/80
PRODUCT V-GRAD Kit
 Product Number V-GRD-KIT
 Expiry Date: 30.11.2017

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,45 - 7,75	7,53
pH (V-GRAD 80)	7,50 - 7,80	7,60
OSMOLALITY (V-GRAD 40)	277-297 mOsm/kg	294 mOsm/kg
OSMOLALITY (V-GRAD 80)	277-297 mOsm/kg	293 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED
MEA 1-cell to expanded blactocyst in 96h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 08.06.2016

