

## CERTIFICATE OF ANALYSIS

**Lot Number:** 605M3601  
**PRODUCT** V-HYLASE  
Product Number V-HYA  
Expiry Date: 31.05.2017

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,40	7,24
OSMOLALITY	257-273 mOsm/kg	267 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,153 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 01.06.2016

