



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 6060115
PRODUCT V-DENUPET 150
Product Number V-DEN-150
Expiry Date: 01.06.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 27.06.2016