

## CERTIFICATE OF ANALYSIS

**Lot Number:** 6060127  
**PRODUCT** V-DENUPET 275  
Product Number V-DEN-275  
Expiry Date: 01.06.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 27.06.2016

