

## CERTIFICATE OF ANALYSIS

**Lot Number:** 606M1402  
**PRODUCT** V-ONESTEP Culture Medium  
Product Number V-OSM-20  
Expiry Date: 26.09.2016

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,40	7,31
OSMOLALITY	257-273 mOsm/kg	263 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80 %	PASSED
Color: Pink rose	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 13.07.2016

