



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 606M3401
PRODUCT V-PVP 7%
Product Number V-PVP-7
Expiry Date: 30.06.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,45	7,30
OSMOLALITY	280 - 320 mOsm/kg	296 Osm/kg
ENDOTOXIN	< 0,5 EU/mL	< 0,05 EU/mL
MEA expanded blastocyste after 96h	≥ 80%	PASSED
STERILITY	PASSED	PASSED
HSSA	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 01.07.2016