

CERTIFICATE OF ANALYSIS

Lot Number: 6070114
PRODUCT V-DENUPET 140
Product Number V-DEN-140
Expiry Date: 11.07.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 19.07.2016

