

CERTIFICATE OF ANALYSIS

Lot Number: 6070160
PRODUCT V-DENUPET 600
Product Number V-DEN-600
Expiry Date: 11.07.2019

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| Gamma Sterilization | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blactocyst in 96 h | ≥ 80 % | PASSED |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 19.07.2016

