

## CERTIFICATE OF ANALYSIS

---

**Lot Number:** 6080112  
**PRODUCT** V-DENUPET 125  
Product Number V-DEN-125  
Expiry Date: 08.08.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Anja Koch ON 22.08.2016

