

CERTIFICATE OF ANALYSIS

Lot Number: 6080130
PRODUCT V-DENUPET 300
Product Number V-DEN-300
Expiry Date: 08.08.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Anja Koch ON 22.08.2016


