



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 6080212/213/214/215/217/220/227/230/260
PRODUCT NAME: V-DENUPET 125, 130, 140, 150, 175, 200, 270, 300, 600
REFERENCE: V-DEN-125, -130, -140, -150, -175, -200, -270, -300, -600
EXPIRY DATE: 25.08.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 21.10.2016