



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 608HO20/HO25/HO30/HO35
PRODUCT NAME: V-ICSIPET HOLDING 20°, 25°, 30°, 35°
REFERENCE: V-HOLD-20, -25, -30, -35
EXPIERY DATE: 29.08.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED
HSMA	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 07.09.2016