

## CERTIFICATE OF ANALYSIS

---

**Lot Number:** 608M0201  
**PRODUCT** V-GRAD 80%  
Product Number V-GRD-80  
Expiry Date: 28.02.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,35
OSMOLALITY	310 - 330 mOsm/kg	316 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	96%

PRODUCT RELEASED BY Monique Slaba ON 05.09.2016

