

CERTIFICATE OF ANALYSIS

Lot Number: 608M0501
PRODUCT V-SPERM WASH
Product Number V-SWM-100
Expiry Date: 31.08.2017

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,25
OSMOLALITY	257-273 mOsm/kg	267 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 15.09.2016

