



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

**Lot Number:** 608M1401  
**PRODUCT** V-ONESTEP Culture Medium  
Product Number V-OSM-20  
Expiry Date: 20.11.2016

| TEST                                      | SPECIFICATION   | RESULT       |
|---|-----------------|--------------|
| pH  | 7,20 - 7,40     | 7,36         |
| OSMOLALITY                                | 257-273 mOsm/kg | 264 mOsm/kg  |
| BACTERIAL SCREEN                          | PASSED          | PASSED       |
| FUNGAL SCREEN                             | PASSED          | PASSED       |
| ENDOTOXIN                                 | < 0,25 EU/mL    | < 0,05 EU/mL |
| MEA 1-cell to expanded blactocyst in 120h | ≥ 80%           | PASSED       |
| Color: Pink rose                          | PASSED          | PASSED       |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 07.09.2016