

CERTIFICATE OF ANALYSIS

Lot Number: 608M3601
PRODUCT V-HYLASE
Product Number V-HYA
Expiry Date: 31.08.2017

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,40	7,21
OSMOLALITY	257-273 mOsm/kg	268 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,2 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Monique Slaba ON 24.08.2016

