

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 6090112/113/114/115/117/120/127/130/160  
**PRODUCT NAME:** V-DENUPET 125, 135, 140, 150, 175, 200, 275, 300, 600  
**REFERENCE:** V-DEN-125, -135, -140, -150, -175, -200, -275, -300, -600  
**EXPIERY DATE:** 30.09.2019

| TEST                                      | SPECIFICATION | RESULT |
|---|---------------|--------|
| Gamma Sterilization                       | 25 - 45 kGy   | PASSED |
| ENDOTOXIN                                 | < 0,05 EU/mL  | PASSED |
| MEA 1-cell to expanded blactocyst in 96 h | ≥ 80 %        | PASSED |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 08.11.2016

