

CERTIFICATE OF ANALYSIS

Lot Number: 609M1401
PRODUCT V-ONESTEP Culture Medium
 Product Number V-OSM-20
 Expiry Date: 18.12.2016

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,45	7,35
OSMOLALITY	257-273 mOsm/kg	262 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80%	PASSED
SMA - Sperm Motility Assay	PASSED	PASSED
Color: Pink rose	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 04.10.2016

