

CERTIFICATE OF ANALYSIS

Lot Number: 609M2302
PRODUCT V-HEPES PLUS 100
Product Number V-HTF-P1
Expiry Date: 30.09.2017

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,17
OSMOLALITY	257-273 mOsm/kg	265 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80%	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 29.09.2016

