

CERTIFICATE OF ANALYSIS

Lot Number: 609M3601
PRODUCT V-HYLASE
Product Number V-HYA
Expiry Date: 30.09.2017

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,40	7,21
OSMOLALITY	257-273 mOsm/kg	267 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,2 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80%	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 27.09.2016

