

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 6100112/114/115/117/120/127/130  
**PRODUCT NAME:** V-DENUPET 125, 140, 150, 175, 200, 275, 300  
**REFERENCE:** V-DEN-125, -140, -150, -175, -200, -275, -300  
**EXPIERY DATE:** 28.10.2019

| TEST                                      | SPECIFICATION | RESULT |
|---|---------------|--------|
| Gamma Sterilization                       | 25 - 45 kGy   | PASSED |
| ENDOTOXIN                                 | < 0,05 EU/mL  | PASSED |
| MEA 1-cell to expanded blactocyst in 96 h | ≥ 80 %        | PASSED |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 09.12.2016

