

## CERTIFICATE OF ANALYSIS

---

Lot Number: 610M0401  
 PRODUCT V-GRAD Kit  
 Product Number V-GRD-KIT  
 Expiry Date: 30.04.2018

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20 - 7,60	7,37
pH (V-GRAD 80)	7,20 - 7,60	7,40
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	321 mOsm/kg
OSMOLALITY (V-GRAD 80)	310 - 330 mOsm/kg	324 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 08.11.2016

